

Analytische Qualitätssicherung Baden-Württemberg

Proficiency Tests UKWIR 10/21 to 14/21
priority substances in surface water -

cypermethrin, hexabromocyclododecanes, nonylphenol, octylphenol,
triclosan, tributyltin

Stuttgart, December 2021

provided by
AQS Baden-Württemberg
c/o Institute for Sanitary Engineering, Water Quality and Solid Waste Management,
University of Stuttgart
Bandtäle 2, 70569 Stuttgart-Büsnau, Germany

Responsibilities:

Scientific director AQS: Dr.-Ing. Dipl.-Chem. Michael Koch
PT manager: Dr.-Ing. Frank Baumeister
Sample preparation: Gertrud Joas, Cornelia Orth

Release of the report: 01 December 2021
by Scientific director AQS: Dr.-Ing. Michael Koch

AQS Baden-Württemberg at
Institute of Sanitary Engineering,
Water Quality and Solid Waste Management
at University of Stuttgart
Bandtäle 2
70569 Stuttgart-Büsnaus
Germany
<http://www.aqsbw.de>
Tel.: +49 (0)711 / 685-65446
Fax: +49 (0)711 / 685-53769
E-Mail: info@aqsbw.de

Table of contents

General	1
PT design	1
Analytical methods.....	2
Evaluation procedure	2
Results of evaluation	3
Explanation of tables and graphs in the appendix.....	3

Appendix A

cypermethrin	A-1
hexabromocyclododecanes	A-8
nonylphenol.....	A-15
octylphenol.....	A-25
triclosan	A-35
tributyltin.....	A-45

General

These PT were provided by AQS Baden-Württemberg as a subcontractor for wca environment limited, Brunel House, Volunteer Way, Faringdon, Oxfordshire SN7 7YR, United Kingdom on behalf of United Kingdom Water Industry Research (UKWIR).

In five rounds the following determinands were to be measured:

- UKWIR 10/21
 - cypermethrin
- UKWIR 11/21
 - hexabromocyclododecanes
- UKWIR 12/21
 - nonylphenol
 - octylphenol
- UKWIR 13/21
 - triclosan
- UKWIR 14/21
 - tributyltin

The PTs were executed and evaluated according to the requirements of ISO 13528: 2015.

PT design

Each participant received the following samples for the PT rounds:

- 3 spiked samples for the determination of the respective determinands based on a filtered surface water (for UKWIR 12/21 waste water treatment plant effluent and for UKWIR 14/21 drinking water) sample in 1000 ml glass bottles with ground glass plug.
- 1 blank sample of the filtered surface water, waste water treatment plant effluent or drinking water resp.

The concentrations of the analytes of the spiked samples were chosen according to the requirements of UKWIR based on the European Regulation for Environmental Quality Standard (Directive 2008/105/EG on environmental quality standards in the field of water policy).

The samples were cooled directly after preparation and dispatched with freezer packs added to the packages by express service (GoExpress). Participants were requested to start with the analysis one day after receipt of the samples at the latest.

Analytical methods

The participants were free to choose a suitable method, but the following limits of quantification were required:

Determinand	Required LOQ
cypermethrin	0.4 ng/l
hexabromocyclododecanes	9.2 ng/l
nonylphenol	50 ng/l
octylphenol	10 ng/l
triclosan	10 ng/l
tributyltin	0.12 ng/l

The samples had to be analysed in duplicate over the complete method (sample preparation and measurement). The participants were asked to submit the results as average values with three significant digits.

Evaluation procedure

The statistical evaluation was executed according to ISO 13528:2015.

Consensus means were calculated for the UKWIR 10/21 and UKWIR 11/21 PTs as well as for UKWIR 13/21 and UKWIR 14/21. For UKWIR 12/21 (nonylphenol and octylphenol) the sum of the spiked amount and the matrix content was used. The matrix content was determined from the median of the matrix measurements of the participants.

The uncertainty of the assigned value was calculated according to ISO 13528:

$$u(x_{pt}) = \frac{1,25 \times s^*}{\sqrt{p}}$$

s*: standard deviation calculated from the results using the Q-method

p: number of results

For UKWIR 12/21 the uncertainty was combined from the uncertainty of the median of the matrix measurements and the uncertainty budget of the spiking procedure.

The standard deviation for proficiency assessment σ_{pt} was calculated in accordance with the European QA/QC Directive: $\sigma_{pt} = 0,25 * x_{pt}$.

A z-score was calculated for each measurement result derived from the assigned value x_{pt} and the standard deviation for proficiency assessment σ_{pt} :

$$z = \frac{x - x_{pt}}{\sigma_{pt}}$$

The assessment of the results was as follows:


- | | |
|-------------------|----------------|
| $ z \leq 2.0$ | satisfactory |
| $2.0 < z < 3.0$ | questionable |
| $ z \geq 3.0$ | unsatisfactory |

Results of evaluation

Number of participants:

PT round	determinands	number of participants	number of participants reporting results
UKWIR 10/21	Cypermethrin	4	4
UKWIR 11/21	hexabromocyclododecanes	6	6
UKWIR 12/21	nonylphenol octylphenol	12	11
UKWIR 13/21	triclosan	10	9
UKWIR 14/21	tributyltin	6	6

In the following figure the percentage of satisfactory, questionable and unsatisfactory results are illustrated.



Explanation of tables and graphs in the appendix

The appendix contains the PT data for all parameters and all samples in tables and graphs. For each parameter the following illustrations are given:

Parameter table

In these tables the following values for each concentration level are listed:

- assigned value in ng/l
- expanded uncertainty of the assigned value in %
- standard deviation of the data set in ng/l, calculated using the Q-method (due to the often low number of participants this standard deviation estimate is not very reliable)

- standard deviation for proficiency assessment in ng/l for the calculation of z-scores (25 % of the assigned value)
- rel. standard deviation for proficiency assessment in %
- tolerance limits above and below in ng/l and % (limit for assessment as 'satisfactory')
- number of values in this level
- number of not satisfactory values ('questionable' or unsatisfactory) below and above the assigned value and the percentage of these values in total

Relative standard deviation

The diagrams for the rel. standard deviation vs. the assigned value show the values compared to the fixed standard deviation for proficiency assessment (horizontal line at 25%) and the concentration dependence.

Sample table

In this table all results of the participants are noted together with uncertainties (where reported). For these uncertainties ζ -scores (zeta-scores) are calculated according to the formula

$$\zeta = \frac{x - x_{pt}}{\sqrt{u_{lab}^2 + u_{x_{pt}}^2}}$$

with

x = result of the participant

x_{pt} = assigned value

u_{lab} = participant's standard uncertainty

$u_{x_{pt}}$ = standard uncertainty of the assigned value

ζ -scores can be used for the plausibility check of measurement uncertainties. The type of assessment is equivalent to that of z-scores, i.e. an absolute value of $\leq 2,0$ can be regarded as 'satisfactory'.

ζ -scores above this value indicate an underestimation of the measurement uncertainty. This table also contains the assigned value and its uncertainty as well as the tolerance limits ($z = \pm 2,0$).

Sample graphs of concentrations

All participants' results, sorted for values, are shown here versus the laboratory codes. The assigned value and its uncertainty as well as the tolerance limits are also included.

z-score graphs

In a similar way the z-scores attributed to the participants' results are shown here versus the laboratory codes.

Graphs of expanded uncertainty


The expanded uncertainty is shown if laboratories reported uncertainties.

ζ -score graphs

If laboratories reported uncertainties, ζ -scores were calculated and are shown versus the laboratory codes.

UKWIR 10/21**cypermethrin**

level	assigned value [ng/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [ng/l]	standard deviation for proficiency assessment [ng/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [ng/l]	lower tolerance limit [ng/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	1,149	130,40	1,198	0,2872	25	1,723	0,5744	50	-50	4	1	1	50,0
2	1,053	12,90	0,1087	0,2633	25	1,580	0,5267	50	-50	4	0	0	0,0
3	1,276	60,44	0,6169	0,3190	25	1,914	0,6380	50	-50	4	0	0	0,0
								sum		12	1	1	16,7


Relative standard deviation


25 % is the value used as standard deviation for proficiency assessment.

UKWIR PT 10/21		cypermethrin - 1		
assigned value [ng/l]*		1,149	± 1,498	
upper tolerance limit [ng/l]		1,723		
lower tolerance limit [ng/l]		0,5744		
lab code	result [ng/l]	±	z-score	assessm.**
1	1,356		0,7	s
2	1,97		2,9	q
3	0,816		-1,2	s
7	0,453		-2,4	q

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor k=2 corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory







UKWIR PT 10/21		cypermethrin - 2		
assigned value [ng/l]*		1,053	± 0,136	
upper tolerance limit [ng/l]		1,58		
lower tolerance limit [ng/l]		0,5267		
lab code	result [ng/l]	±	z-score	assessm.**
1	1,019		-0,1	s
2	1,1		0,2	s
3	1,068		0,1	s
7	0,591		-1,8	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor k=2 corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory







UKWIR PT 10/21		cypermethrin - 3			
assigned value [ng/l]*			1,276	± 0,771	
upper tolerance limit [ng/l]			1,914		
lower tolerance limit [ng/l]			0,638		
lab code	result [ng/l]	±	ζ-score	z-score	assessm.**
1	1,076			-0,6	s
2	1,75			1,5	s
3	1,354			0,2	s
7	0,924			-1,1	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor k=2 corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory






UKWIR 11/21

Hexabromocyclododecane

level	assigned value [ng/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [ng/l]	standard deviation for proficiency assessment [ng/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [ng/l]	lower tolerance limit [ng/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	18,59	25,56	4,655	4,647	25,00	27,88	9,294	50,00	-50,00	6	0	0	0,0
2	33,03	19,23	6,222	8,258	25,00	49,55	16,52	50,00	-50,00	6	0	0	0,0
3	46,79	17,27	7,918	11,70	25,00	70,19	23,40	50,00	-50,00	6	0	0	0,0
								sum	sum	18	0	0	0,0


Relative standard deviation


25 % is the value used as standard deviation for proficiency assessment.

UKWIR PT 11/21		Hexabromocyclododecane - 1			
lab code	result [ng/l]	±	ζ -score	z-score	assessm.**
1	22,9			0,9	s
7	15,2			-0,7	s
9	17,8			-0,2	s
10	14,4			-0,9	s
13	21,5			0,6	s
15	19,73			0,2	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor k=2 corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory







UKWIR PT 11/21		Hexabromocyclododecane - 2			
assigned value [ng/l]*		33,03	± 6,35		
upper tolerance limit [ng/l]		49,55			
lower tolerance limit [ng/l]		16,52			
lab code	result [ng/l]	±	ζ -score	z-score	assessm.**
1	36			0,4	s
7	34,2			0,1	s
9	31,6			-0,2	s
10	18,9			-1,7	s
13	40,1			0,9	s
15	32,585			-0,1	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor k=2 corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory







UKWIR PT 11/21		Hexabromocyclododecane - 3			
assigned value [ng/l]*		46,79	± 8,08		
upper tolerance limit [ng/l]		70,19			
lower tolerance limit [ng/l]		23,4			
lab code	result [ng/l]	±	ζ-score	z-score	assessm.**
1	46,5			0,0	s
7	48,5			0,1	s
9	48,2			0,1	s
10	38			-0,8	s
13	59,9			1,1	s
15	40,886			-0,5	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor k=2 corresponding to a confidence level of about 95%


** s = satisfactory, q = questionable, u = unsatisfactory





UKWIR 12/21:

Nonylphenol													
level	assigned value [ng/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [ng/l]	standard deviation for proficiency assessment [ng/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [ng/l]	lower tolerance limit [ng/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	635,4	15,59	135,0	158,9	25	953,2	317,7	50	-50	8	1	1	20,0
2	1154	8,65	453,8	288,5	25	1731	576,9	50	-50	9	1	0	10,0
3	1497	6,74	383,9	374,1	25	2245	748,3	50	-50	9	1	0	10,0
								sum		26	3	1	15,4


Relative standard deviation



25 % is used as standard deviation for proficiency assessment.


UKWIR PT 12/21		Nonylphenol - 1			
lab code	result [ng/l]	±	ζ -score	z-score	assessm.**
1	600			-0,2	s
3	610			-0,2	s
4	692,7			0,4	s
5	377			-1,6	s
8	35,8	3,58	-12,1	-3,8	u
11	645	10	0,2	0,1	s
16	400			-1,5	s
17	881650			5545,9	u
18	684			0,3	s
20	812			1,1	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor k=2 corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory











UKWIR PT 12/21		Nonylphenol - 2			
lab code	result [ng/l]	±	ζ -score	z-score	assessm.**
1	1060			-0,3	s
3	940			-0,7	s
4	1323,75			0,6	s
5	721			-1,5	s
8	30,99	3,1	-22,5	-3,9	u
11	1080	10	-1,5	-0,3	s
16	750			-1,4	s
17	1460			1,1	s
18	1077			-0,3	s
20	1820			2,3	q

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor k=2 corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory











UKWIR PT 12/21		Nonylphenol - 3			
assigned value [ng/l]*		1497	± 101		
lab code	result [ng/l]	±	ζ -score	z-score	assessm.**
1	1340			-0,4	s
3	1250			-0,7	s
4	2081,75			1,6	s
5	1138			-1,0	s
8	28,11	3	-29,1	-3,9	u
11	1420	10	-1,5	-0,2	s
16	980			-1,4	s
17	1550			0,1	s
18	1374			-0,3	s
20	1810			0,8	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor k=2 corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory








Octylphenol


level	assigned value [ng/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [ng/l]	standard deviation for proficiency assessment [ng/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [ng/l]	lower tolerance limit [ng/l]	upper tolerance limit [%]	number of results	out below	out above	out [%]
1	43,85	7,59	15,64	10,96	25	65,78	21,93	50	-50	9	1	2 27,3
2	131,0	3,07	45,19	32,75	25	196,5	65,50	50	-50	10	1	0 9,1
3	219,7	2,40	46,08	54,92	25	329,5	109,8	50	-50	10	2	0 18,2
								sum	29	4	2	20,7






UKWIR PT 12/21		Octylphenol - 1			
assigned value [ng/l]*		43,85	± 3,33		
lab code	result [ng/l]	±	ζ -score	z-score	assessm.**
1	46,1			0,2	s
3	50			0,6	s
4	38,85			-0,5	s
5	24			-1,8	s
8	3,35	0,4	-24,2	-3,7	u
11	48,4	3,9	1,8	0,4	s
16	32			-1,1	s
17	68014			6199,9	u
18	50,1			0,6	s
20	93,1			4,5	u
21	40,6			-0,3	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor k=2 corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory











UKWIR PT 12/21		Octylphenol - 2			
lab code	result [ng/l]	±	ζ -score	z-score	assessm.**
1	127			-0,1	s
3	130			0,0	s
4	102,45			-0,9	s
5	78			-1,6	s
8	5,7	0,68	-61,5	-3,8	u
11	130	3,9	-0,4	0,0	s
16	91			-1,2	s
17	159			0,9	s
18	123			-0,2	s
20	198			2,0	s
21	104			-0,8	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor k=2 corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory











UKWIR PT 12/21		Octylphenol - 3			
lab code	result [ng/l]	±	ζ -score	z-score	assessm.**
1	210			-0,2	s
3	210			-0,2	s
4	164,75			-1,0	s
5	114			-1,9	s
8	4,99	0,6	-80,9	-3,9	u
11	207	3,9	-3,9	-0,2	s
16	154			-1,2	s
17	222			0,0	s
18	198,4			-0,4	s
20	299			1,4	s
21	177			-0,8	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor k=2 corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory








UKWIR 13/21:

triclosan														
	level	assigned value [ng/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [ng/l]	standard deviation for proficiency assessment [ng/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [ng/l]	lower tolerance limit [ng/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	29,98	20,86	7,505	7,496	25	44,97	14,99	50	-50	9	0	0	0,0	
2	72,39	26,44	22,97	18,10	25	108,6	36,19	50	-50	9	0	0	0,0	
3	114,3	14,73	20,19	28,57	25	171,4	57,14	50	-50	9	0	0	0,0	
									sum	27	0	0	0,0	


Relative standard deviation



25 % is used as standard deviation for proficiency assessment.


UKWIR PT 13/21		triclosan - 1			
	assigned value [ng/l]*	29,98	± 6,25		
	upper tolerance limit [ng/l]	44,97			
	lower tolerance limit [ng/l]	14,99			
lab code	result [ng/l]	±	ζ-score	z-score	assessm.**
1	38			1,1	s
2	29,2	11,7	-0,1	-0,1	s
11	36,45			0,9	s
13	29,5	2,9	-0,1	-0,1	s
14	29,2	12,5	-0,1	-0,1	s
15	36,3			0,8	s
16	20			-1,3	s
17	25,7			-0,6	s
19	25,5	5,47	-1,1	-0,6	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor k=2 corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory











UKWIR PT 13/21		triclosan - 2			
assigned value [ng/l]*		72,39	\pm 19,14		
upper tolerance limit [ng/l]		108,6			
lower tolerance limit [ng/l]		36,19			
lab code	result [ng/l]	\pm	ζ -score	z-score	assessm.**
1	92			1,1	s
2	62,5	25	-0,6	-0,5	s
11	76,6			0,2	s
13	80,6	2,9	0,8	0,5	s
14	62,6	26,9	-0,6	-0,5	s
15	89,3			0,9	s
16	80			0,4	s
17	61,5			-0,6	s
19	46,4	9,98	-2,4	-1,4	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor k=2 corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory











UKWIR PT 13/21		triclosan - 3			
lab code	result [ng/l]	±	ζ -score	z-score	assessm.**
1	143			1,0	s
2	85,8	34,3	-1,5	-1,0	s
11	118,2			0,1	s
13	110	2,9	-0,5	-0,1	s
14	118	50,7	0,1	0,1	s
15	147			1,1	s
16	100			-0,5	s
17	105,9			-0,3	s
19	103	22,1	-0,8	-0,4	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor k=2 corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory








UKWIR 14/21:

Tributyltin														
	level	assigned value [ng/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [ng/l]	standard deviation for proficiency assessment [ng/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [ng/l]	lower tolerance limit [ng/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	0,7287	30,91	0,2207	0,1822	25	1,093	0,3644	50	-50	6	0	0	0,0	
2	1,039	45,79	0,4660	0,2597	25	1,558	0,5193	50	-50	6	0	0	0,0	
3	1,306	19,31	0,2471	0,3266	25	1,960	0,6532	50	-50	6	0	0	0,0	
								sum	18	0	0	0	0,0	


Relative standard deviation



25 % is used as standard deviation for proficiency assessment.


UKWIR PT 14/21		Tributyltin - 1			
lab code	result [ng/l]	\pm	ζ -score	z-score	assessm.**
2	0,691	0,229	-0,2	-0,2	s
3	0,77			0,2	s
5	0,55			-1,0	s
12	0,668			-0,3	s
13	0,896			0,9	s
14	0,7974			0,4	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor k=2 corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory











UKWIR PT 14/21		Tributyltin - 2			
assigned value [ng/l]*		1,039	± 0,476		
upper tolerance limit [ng/l]		1,558			
lower tolerance limit [ng/l]		0,5193			
lab code	result [ng/l]	±	ζ-score	z-score	assessm.**
2	0,879	0,292	-0,6	-0,6	s
3	0,89			-0,6	s
5	0,6			-1,7	s
12	1,56			2,0	s
13	1,058			0,1	s
14	1,245			0,8	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor k=2 corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory







UKWIR PT 14/21		Tributyltin - 3			
assigned value [ng/l]*		1,306	± 0,252		
upper tolerance limit [ng/l]		1,96			
lower tolerance limit [ng/l]		0,6532			
lab code	result [ng/l]	±	ζ-score	z-score	assessm.**
2	1,25	0,415	-0,2	-0,2	s
3	1,2			-0,3	s
5	1,2			-0,3	s
12	1,93			1,9	s
13	1,082			-0,7	s
14	1,4292			0,4	s

* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor k=2 corresponding to a confidence level of about 95%

** s = satisfactory, q = questionable, u = unsatisfactory

