

**University of Stuttgart**  
Germany



# Analytische Qualitätssicherung Baden-Württemberg

---

Proficiency Test 7/23  
- TW S11 – Haloacetic acids in drinking water -

## Final report

provided by  
AQS Baden-Württemberg at  
Institute for Sanitary Engineering, Water Quality and Solid Waste Management,  
University of Stuttgart  
Bandtäle 2, 70569 Stuttgart-Büsnau, Germany



And  
IWW Water Center  
Moritzstr. 26, 45476 Mülheim an der Ruhr, Germany



Stuttgart, in January 2024

**AQS Baden-Württemberg at  
Institute of Sanitary Engineering,  
Water Quality and Solid Waste Management  
at University of Stuttgart  
Bandtäle 2  
70569 Stuttgart-Büsnau  
Germany  
<http://www.aqsbw.de>  
Tel.: +49 (0)711 / 685-65446  
Fax: +49 (0)711 / 685-53769  
E-Mail: [info@aqsbw.de](mailto:info@aqsbw.de)**

**Responsibilities:**

<b>Scientific director:</b>	Dr.-Ing. Michael Koch	
<b>PT coordinator:</b>	Dr.-Ing. Frank Baumeister	
<b>Assistant PT coordinator</b>	Dipl.-Biol. Biljana Marić	
<b>Sample preparation</b>	Dr. Vasil Valkov (IWW)	
<b>Release of the report:</b>	Dr.-Ing. Michael Koch	on 9.1.24
<b>Version of the report</b>	1	

## List of contents

1. General .....	1
2. PT design .....	1
3. Sample preparation .....	1
4. Sample distribution.....	1
5. Analytical methods .....	1
6. Submission of the results .....	2
7. Basic principle of evaluation and assessment.....	2
8. Evaluation.....	3
9. Explanation for the appendices .....	3
10. Measurement uncertainty.....	4
11. Note to monobromo acetic acid .....	4
12. Traceable reference values .....	5
13. Internet.....	5

### Appendix A

MONOCHLOROACETIC ACID .....	A-1
DICHLOROACETIC ACID .....	A-7
TRICHLOROACETIC ACID.....	A-13
DIBROMOACETIC ACID .....	A-19

### Appendix B

### Appendix C

MONOCHLOROACETIC ACID .....	C-1
DICHLOROACETIC ACID .....	C-10
TRICHLOROACETIC ACID.....	C-19
DIBROMOACETIC ACID .....	C-28

## 1. General

This PT was provided by AQS Baden-Württemberg in cooperation with IWW Water Center in Mülheim an der Ruhr and with the network “NORMAN” (Network of reference laboratories for monitoring of emerging environmental pollutants).

The following parameters were offered:

- monochloroacetic acid
- dichloroacetic acid
- trichloroacetic acid
- monobromoacetic acid
- dibromoacetic acid

The PT was executed and evaluated according to the requirements of DIN 38402-A45 and ISO/TS 20612.

## 2. PT design

Each participant received the following samples:

- 3 samples for the determination of the mentioned parameter in 1000-ml ground glass bottles (brown) with ground glass plug. Preservation by adding hydrochloric acid (pH 2) and sodium thiosulfate.

3 different concentration levels/batches were produced. All participants received the same samples.

## 3. Sample preparation

The samples for the determination of the haloacetic acids were based on a real ground water matrix. The ground water was used without treatment for the sample preparation.

The ground water was spiked with stock solutions and the concentrations covered drinking and ground water relevant ranges.

## 4. Sample distribution

The samples were dispatched on 15.08.2023 by express service.

## 5. Analytical methods

The participants were free to choose a suitable method, but a limit of quantification of 1 µg/l for all parameters was required.

The participants were informed that the samples had to be analysed in the own laboratory, with own personal and own equipment. Subcontracting of the analysis was not allowed.

The participants were informed to cool the samples after receipt and to start with the analysis one day after receipt at the latest.

The samples had to be analysed in duplicate over the complete method (sample preparation and measurement). The participants were asked to report the results as average means from both determinations in  $\mu\text{g/l}$  with three significant digits.

## 6. Submission of the results

The deadline for the submission of results was on 04.09.2023.

## 7. Basic principle of evaluation and assessment

The basic principle of the evaluation and assessment of the PTs from AQS Baden-Württemberg are described in the document „Evaluation of the PTs and information for the report“, which can be downloaded from [www.aqsbw.de/pdf/ausw\\_berichte\\_v1\\_en.pdf](http://www.aqsbw.de/pdf/ausw_berichte_v1_en.pdf).


This PT was evaluated as follows:

<b>Assigned value <math>x_{\text{pt}}</math>:</b>	Consensus value (Hampel estimator)						
<b>Standard deviation for proficiency assessment <math>\sigma_{\text{pt}}</math>:</b>	Q method						
<b>Upper limit of <math>\sigma_{\text{pt}}</math>:</b>	25 %						
<b>Lower limit of <math>\sigma_{\text{pt}}</math>:</b>	5 %						
<b>Assessment:</b>	$z_U$ -Score						
<b>Classification of the single results:</b>	<table style="border: none;"> <tr> <td><math> z_U  \leq 2.0</math></td> <td>successful</td> </tr> <tr> <td><math>2.0 &lt;  z_U  &lt; 3.0</math></td> <td>questionable</td> </tr> <tr> <td><math> z_U  \geq 3.0</math></td> <td>unsatisfactory</td> </tr> </table>	$ z_U  \leq 2.0$	successful	$2.0 <  z_U  < 3.0$	questionable	$ z_U  \geq 3.0$	unsatisfactory
$ z_U  \leq 2.0$	successful						
$2.0 <  z_U  < 3.0$	questionable						
$ z_U  \geq 3.0$	unsatisfactory						
<b>Parameter assessment:</b>	A parameter was assessed as successful, if more than half of the values were correctly determined (2 out of 3 values are within the tolerance limits).						

## 8. Evaluation

<b>Number of participants:</b>	39
<b>Number of reported values</b>	350
<b>Number of accepted values:</b>	293 (83,7 %)

**Illustration of the successful and not successful laboratories for each parameter**



The parameter monobromoacetic acid could not be evaluated. Explanation is given in chapter 12 “Note to monobromoacetic acid”.

## 9. Explanation for the appendices

The explanations for the appendices can be found in the document „Evaluation of the PTs and information for the report“, which can be downloaded from [www.aqsbw.de/pdf/ausw\\_berichte\\_v1\\_en.pdf](http://www.aqsbw.de/pdf/ausw_berichte_v1_en.pdf).

## 10. Measurement uncertainty

### General:

Number of labs with valid values	32
Number of labs with valid values and reported measurement uncertainties	16 (50 %)
Number of valid values	362
Number of valid values with measurement uncertainties	185 (51,1 %)

### Measurement uncertainties against the accreditation status

Accreditation status of the values	Number of values	Number of values with measurement uncertainty
accredited	105	80 (76,2 %)
not accredited	162	81 (50 %)
not specified	83	24 (28,9 %)

### Interpretation of the reported measurement uncertainties:

If measurement uncertainties are underestimated values assessed as “satisfactory” in the PT ( $|z_U| \leq 2$ ), will have a large  $\zeta$ -score.  $|\zeta| > 2$  means that the “own” requirements (defined in terms of estimated uncertainty) are not fulfilled.

<b>Number of values with reported measurement uncertainty having a <math> z_U  \leq 2,0</math></b>	156
<b>Number of values with a magnitude of <math>\zeta</math>-scores <math>&gt; 2</math></b> The own requirements of the laboratory are not fulfilled and the estimation of the measurement uncertainty is too low	34 (21,8 %)

## 11. Note to monobromo acetic acid

The monobromoacetic acid was no longer detectable in the samples due to an error during sample stabilisation. Sodium thiosulphate was added to the samples as required by the standard. In this case, however, the standard specifications only allow a period of 3 days until analysis, which is too short. With the intention of stabilising the samples even longer, they were also acidified, but this, together with the thiosulphate, led to the monobromoacetic acid breaking down after a very short time and no longer being detectable. This point had not been observed in the preliminary tests.

The acidification apparently had no influence on the other parameters.

## 12. Traceable reference values

The explanations about traceable reference values can be found in the document „Evaluation of the PTs and information for the report“, which can be downloaded from [www.aqsbw.de/pdf/ausw\\_berichte\\_v1\\_en.pdf](http://www.aqsbw.de/pdf/ausw_berichte_v1_en.pdf)

## 13. Internet


The report is available on the following webpage: [www.aqsbw/pdf/283/report\\_283.pdf](http://www.aqsbw/pdf/283/report_283.pdf)



# monochloroacetic acid


level	assigned value [µg/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [µg/l]	standard deviation for proficiency assessment [µg/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [µg/l]	lower tolerance limit [µg/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	4,264	13,03	1,176	1,066	25,00	6,736	2,337	57,99	-45,19	28	4	0	14,3
2	16,54	9,82	3,438	3,438	20,79	24,29	10,24	46,87	-38,10	28	6	0	20,7
3	42,92	8,19	7,440	7,440	17,33	59,32	29,12	38,23	-32,16	28	8	0	27,6
sum										84	18	0	21,4

## Recovery and matrix content




slope of the regression: 0,854; average recovery rate: 85,4 %  
 neg. x-intercept corresponds to the matrix content: 0,0046 µg/l  
 exp. Uncertainty of the matrix content: 0,0046 µg/l = 100 %



Relative standard deviation and tolerance limits



The relative standard deviations calculated with the Q-method reached the upper limit with one concentration level.




Method specific evaluation






The values determined with IC-MS-direct injection showed the closest statistical distribution. The values determined with LC-MS showed a higher ratio of to high values.

**Comparison of means and reference values**

level	mean [µg/l]	exp. uncertainty [µg/l]	exp. uncertainty [%]	reference value [µg/l]	exp. uncertainty [µg/l]	exp. uncertainty [%]
1	4,264	0,556	13,0	4,957	0,090	1,8
2	16,54	1,62	9,8	19,82	0,36	1,8
3	42,92	3,51	8,2	49,53	0,89	1,8








LC-MS - direct injection									
level	robust mean [µg/l]	exp. unc. of the mean [µg/l]	exp. unc. of the mean [%]	robust standard deviation [µg/l]	robust standard deviation [%]	number of results	out below	out above	out [%]
1	4,28	0,344	8,037	1,232	28,76	20	2	0	10
2	15,7	0,996	6,334	3,65	23,22	21	2	0	9,5238
3	39,8	2,333	5,857	8,552	21,47	21	3	0	14,286

# dichloroacetic acid

level	assigned value [µg/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [µg/l]	standard deviation for proficiency assessment [µg/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [µg/l]	lower tolerance limit [µg/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	4,436	19,35	1,849	1,109	25,00	7,008	2,431	57,99	-45,19	29	4	2	20,0
2	17,47	13,48	5,071	4,366	25,00	27,59	9,574	57,99	-45,19	29	4	2	20,0
3	45,83	14,28	14,09	11,46	25,00	72,40	25,12	57,99	-45,19	29	3	0	10,0
sum										87	11	4	17,2

## Recovery and matrix content




slope of the regression: 0,881; average recovery rate: 88,1 %


neg. x-intercept corresponds to the matrix content: 0 µg/l

exp. Uncertainty of the matrix content: 0,1 µg/l = 0 %

Relative standard deviation and tolerance limits





The relative standard deviations calculated with the Q-method reached the upper limit with all concentration levels.






Method specific evaluation






The values determined with LC-MS - direct injection showed the closest statistical distribution.

**Comparison of means and reference values**

level	Experimental Data			Reference Data		
	mean [ $\mu\text{g/l}$ ]	exp. uncertainty [ $\mu\text{g/l}$ ]	exp. uncertainty [%]	reference value [ $\mu\text{g/l}$ ]	exp. uncertainty [ $\mu\text{g/l}$ ]	exp. uncertainty [%]
1	4,436	0,859	19,4	5,094	0,135	2,6
2	17,47	2,35	13,5	20,38	0,37	1,8
3	45,83	6,54	14,3	50,94	0,90	1,8








LC-MS - direct injection									
level	robust mean [µg/l]	exp. unc. of the mean [µg/l]	exp. unc. of the mean [%]	robust standard deviation [µg/l]	robust standard deviation [%]	number of results	out below	out above	out [%]
1	4,3994	0,448	10,18	1,6422	37,327	21	2	1	14,29
2	17,463	1,147	6,57	4,2059	24,084	21	3	1	19,05
3	46,051	3,809	8,27	13,963	30,32	21	3	0	14,29

# trichloroacetic acid

level	assigned value [µg/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [µg/l]	standard deviation for proficiency assessment [µg/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [µg/l]	lower tolerance limit [µg/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	4,756	12,60	1,222	1,189	25,00	7,514	2,607	57,99	-45,19	26	2	3	18,5
2	19,89	11,86	4,992	4,972	25,00	31,42	10,90	57,99	-45,19	28	3	2	17,2
3	49,20	9,91	10,32	10,32	20,98	72,50	30,29	47,37	-38,43	28	4	1	17,2
sum										82	9	6	18,3

## Recovery and matrix content




slope of the regression: 0,916; average recovery rate: 91,6 %


neg. x-intercept corresponds to the matrix content: 0 µg/l

exp. Uncertainty of the matrix content: 0,225 µg/l = 0 %



Relative standard deviation and tolerance limits



The relative standard deviations calculated with the Q-method reached the upper limit with one concentration level.




Method specific evaluation





The values determined with LC-MS - direct injection showed the closest statistical distribution.


**Comparison of means and reference values**

level	mean [µg/l]	exp. uncertainty [µg/l]	exp. uncertainty [%]	reference value [µg/l]	exp. uncertainty [µg/l]	exp. uncertainty [%]
1	4,756	0,599	12,6	5,430	0,262	4,8
2	19,89	2,36	11,9	21,72	0,58	2,7
3	49,20	4,88	9,9	54,30	1,36	2,5










LC-MS - direct injection									
level	robust mean [µg/l]	exp. unc. of the mean [µg/l]	exp. unc. of the mean [%]	robust standard deviation [µg/l]	robust standard deviation [%]	number of results	out below	out above	out [%]
1	4,9855	0,231	4,6338	0,7841	15,728	18	2	3	27,78
2	20,653	1,0753	5,2065	3,8472	18,627	20	2	2	20
3	51,301	2,4795	4,8332	8,8708	17,292	20	2	0	10

# dibromoacetic acid


level	assigned value [µg/l]	expanded uncertainty of the assigned value [%]	standard deviation, calculated using robust statistics [µg/l]	standard deviation for proficiency assessment [µg/l]	standard deviation for proficiency assessment [%]	upper tolerance limit [µg/l]	lower tolerance limit [µg/l]	upper tolerance limit [%]	lower tolerance limit [%]	number of results	out below	out above	out [%]
1	4,986	14,02	1,506	1,246	25,00	7,876	2,733	57,99	-45,19	29	6	0	20,7
2	19,00	13,64	5,677	4,749	25,00	30,01	10,41	57,99	-45,19	30	5	2	23,3
3	46,43	15,02	15,28	11,61	25,00	73,36	25,45	57,99	-45,19	30	3	1	13,3
sum										89	14	3	19,1

## Recovery and matrix content




slope of the regression: 0,841; average recovery rate: 84,1 %  
 neg. x-intercept corresponds to the matrix content: 0,425 µg/l  
 exp. Uncertainty of the matrix content: 0,425 µg/l = 100 %



**Relative standard deviation and tolerance limits**



The relative standard deviations calculated with the Q-method reached the upper limit with all concentration levels.




**Method specific evaluation**






The values determined with LC-MS - direct injection showed the closest statistical distribution.

**Comparison of means and reference values**

level	mean [µg/l]	exp. uncertainty [µg/l]	exp. uncertainty [%]	reference value [µg/l]	exp. uncertainty [µg/l]	exp. uncertainty [%]
1	4,986	0,699	14,0	5,932	0,520	8,8
2	19,00	2,59	13,6	22,45	1,27	5,7
3	46,43	6,97	15,0	55,49	3,03	5,5











LC-MS - direct injection									
level	robust mean [ $\mu\text{g/l}$ ]	exp. unc. of the mean [ $\mu\text{g/l}$ ]	exp. unc. of the mean [%]	robust standard deviation [ $\mu\text{g/l}$ ]	robust standard deviation [%]	number of results	out below	out above	out [%]
1	4,9229	0,384	7,7962	1,3731	27,892	20	2	1	15
2	18,917	1,513	7,9999	5,5479	29,328	21	3	0	14,29
3	45,689	3,826	8,3732	14,025	30,697	21	2	0	9,524









PT 7/23		monochloroacetic acid - 1			
assigned value [ $\mu\text{g/l}$ ]*		4,264 $\pm$ 0,556			
upper tolerance limit [ $\mu\text{g/l}$ ]		6,736			
lower tolerance limit [ $\mu\text{g/l}$ ]		2,337			
lab code	result [ $\mu\text{g/l}$ ]	$\pm$	z-score	Z <sub>U</sub> -score	assessm.**
1	5,14	1	1,5	0,7	s
4	5,16			0,7	s
5	4,96	1,28	1,0	0,6	s
8	4,3	1,7	0,0	0,0	s
9	3,28	0,66	-2,3	-1,0	s
12	4,42			0,1	s
13	1,14			-3,2	u
14	2,4			-1,9	s
15	3,02			-1,3	s
17	4,8	0,672	1,2	0,4	s
19	4,002			-0,3	s
20	4,03			-0,2	s
21	5,39	2,7	0,8	0,9	s
22	2,14	0,15	-7,4	-2,2	q
23	4,43	0,665	0,4	0,1	s
26	5,12			0,7	s
27	4,9	0,58	1,6	0,5	s
28	2,07	0,15	-7,6	-2,3	q
29	4,59			0,3	s
30	5	1,25	1,1	0,6	s
31	5,37	0,85	2,2	0,9	s
35	7,47			2,6	q
37	4,22			0,0	s
39	5,82			1,3	s
40	5,61	1,12	2,2	1,1	s
41	5,94			1,4	s
43	1,94	0,39	-6,8	-2,4	q
46	4,79			0,4	s
47	2,6	0,521	-4,4	-1,7	s



\* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor  $k=2$  corresponding to a confidence level of about 95%

\*\* s = satisfactory, q = questionable, u = unsatisfactory



Strongly deviating values are not shown in the diagram.







PT 7/23		monochloroacetic acid - 2			
assigned value [ $\mu\text{g/l}$ ]*		16,54 $\pm$ 1,62			
upper tolerance limit [ $\mu\text{g/l}$ ]		24,29			
lower tolerance limit [ $\mu\text{g/l}$ ]		10,24			
lab code	result [ $\mu\text{g/l}$ ]	$\pm$	z-score	Z <sub>U</sub> -score	assessm.**
1	18,4	3,7	0,9	0,5	s
4	20,3			1,0	s
5	16,9	5,64	0,1	0,1	s
8	18,5	7,1	0,5	0,5	s
9	10,2	2	-4,9	-2,0	s
12	17,3			0,2	s
13	1,48			-4,8	u
14	9,9			-2,1	q
15	19			0,6	s
16	14,8			-0,6	s
17	20,4	3,71	1,9	1,0	s
19	9,053			-2,4	q
20	17,8			0,3	s
21	17,3	8,65	0,2	0,2	s
22	4,66	0,12	-14,6	-3,8	u
23	18,3	2,75	1,1	0,5	s
26	21,8			1,4	s
27	14,9	1,77	-1,4	-0,5	s
28	14	0,8	-2,8	-0,8	s
29	17,8			0,3	s
30	18,1	4,53	0,6	0,4	s
31	19	3,01	1,4	0,6	s
35	25,8			2,4	q
37	16,21			-0,1	s
39	19,49			0,8	s
40	21,3	4,26	2,1	1,2	s
41	20,3			1,0	s
43	9,51	1,9	-5,6	-2,2	q
46	10,67			-1,9	s
47	9,91	1,98	-5,2	-2,1	q



\* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor  $k=2$  corresponding to a confidence level of about 95%

\*\* s = satisfactory, q = questionable, u = unsatisfactory



Strongly deviating values are not shown in the diagram.






Strongly deviating values are not correctly shown in the diagram.




PT 7/23		monochloroacetic acid - 3			
assigned value [ $\mu\text{g/l}$ ]*		42,92 $\pm$ 3,51			
upper tolerance limit [ $\mu\text{g/l}$ ]		59,32			
lower tolerance limit [ $\mu\text{g/l}$ ]		29,12			
lab code	result [ $\mu\text{g/l}$ ]	$\pm$	z-score	Z <sub>U</sub> -score	assessm.**
1	47,6	9,5	0,9	0,6	s
4	50,1			0,9	s
5	47,6	17,7	0,5	0,6	s
8	39,7	15,3	-0,4	-0,5	s
9	24,3	4,9	-6,2	-2,7	q
12	44,2			0,2	s
13	2,07			-5,9	u
14	25,4			-2,5	q
15	45,1			0,3	s
16	41,5			-0,2	s
17	46	10,5	0,6	0,4	s
19	19,926			-3,3	u
20	41,2			-0,2	s
21	45,5	22,8	0,2	0,3	s
22	8,96	0,46	-19,2	-4,9	u
23	42,8	6,42	0,0	0,0	s
26	53,1			1,2	s
27	39,6	4,71	-1,1	-0,5	s
28	48,2	13,1	0,8	0,6	s
29	46,9			0,5	s
30	48,9	12,3	0,9	0,7	s
31	49,1	7,79	1,4	0,8	s
35	50,4			0,9	s
37	39,21			-0,5	s
39	47,44			0,6	s
40	53,9	10,8	1,9	1,3	s
41	49,9			0,9	s
43	25	5	-5,9	-2,6	q
46	24,97			-2,6	q
47	25,6	5,13	-5,6	-2,5	q


\* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor  $k=2$  corresponding to a confidence level of about 95%

\*\* s = satisfactory, q = questionable, u = unsatisfactory




Strongly deviating values are not shown in the diagram.





Strongly deviating values are not correctly shown in the diagram.




Strongly deviating values are not correctly shown in the diagram.


PT 7/23		dichloroacetic acid - 1			
assigned value [ $\mu\text{g/l}$ ]*		4,436 $\pm$ 0,859			
upper tolerance limit [ $\mu\text{g/l}$ ]		7,008			
lower tolerance limit [ $\mu\text{g/l}$ ]		2,431			
lab code	result [ $\mu\text{g/l}$ ]	$\pm$	z-score	Z <sub>U</sub> -score	assessm.**
1	5,43	1,1	1,4	0,8	s
4	6,28			1,4	s
5	5,66	2,58	0,9	1,0	s
8	5,3	1,2	1,2	0,7	s
9	3,45	0,69	-1,8	-1,0	s
12	2,7			-1,7	s
13	0,283			-4,1	u
14	2,4			-2,0	s
15	1,15			-3,3	u
16	4,28			-0,2	s
17	5,24	1,12	1,1	0,6	s
18	2,09	0,209	-5,3	-2,3	q
19	4,846			0,3	s
20	4,35			-0,1	s
21	3,26	1,63	-1,3	-1,2	s
22	1,8	0,18	-6,0	-2,6	q
23	6,44	0,966	3,1	1,6	s
26	5,7			1,0	s
27	4,15	0,477	-0,6	-0,3	s
28	4,25	0,44	-0,4	-0,2	s
29	3,8			-0,6	s
30	5,6	1,4	1,4	0,9	s
31	5,38	0,76	1,6	0,7	s
35	8,01			2,8	q
37	4,55			0,1	s
39	5,47			0,8	s
40	7,36	2,8	2,0	2,3	q
41	5,84			1,1	s
43	3,81	0,76	-1,1	-0,6	s
46	15,1			8,3	u
47	3,25	0,651	-2,2	-1,2	s



\* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor  $k=2$  corresponding to a confidence level of about 95%

\*\* s = satisfactory, q = questionable, u = unsatisfactory



Strongly deviating values are not shown in the diagram.







PT 7/23		dichloroacetic acid - 2			
assigned value [ $\mu\text{g/l}$ ]*		17,47 $\pm$ 2,35			
upper tolerance limit [ $\mu\text{g/l}$ ]		27,59			
lower tolerance limit [ $\mu\text{g/l}$ ]		9,574			
lab code	result [ $\mu\text{g/l}$ ]	$\pm$	z-score	Z <sub>U</sub> -score	assessm.**
1	21,1	4,2	1,5	0,7	s
4	21,8			0,9	s
5	22	13,8	0,6	0,9	s
8	17,2	3,8	-0,1	-0,1	s
9	10,8	2,2	-4,1	-1,7	s
12	12			-1,4	s
13	1,07			-4,2	u
14	9,7			-2,0	s
15	5,12			-3,1	u
16	15,3			-0,5	s
17	20,4	2,2	1,8	0,6	s
18	9,26	0,926	-6,5	-2,1	q
19	15,565			-0,5	s
20	17,5			0,0	s
21	14	7	-0,9	-0,9	s
22	7,01	0,7	-8,5	-2,6	q
23	21	3,15	1,8	0,7	s
26	21,6			0,8	s
27	16,5	1,9	-0,6	-0,2	s
28	20,1	3,53	1,2	0,5	s
29	16,4			-0,3	s
30	20,6	5,15	1,1	0,6	s
31	20,4	2,87	1,6	0,6	s
35	24,3			1,3	s
37	21,4			0,8	s
39	22,44			1,0	s
40	27,9	10,6	1,9	2,1	q
41	22			0,9	s
43	14,5	2,9	-1,6	-0,8	s
46	36,16			3,7	u
47	15,8	3,17	-0,8	-0,4	s

\* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor  $k=2$  corresponding to a confidence level of about 95%


\*\* s = satisfactory, q = questionable, u = unsatisfactory




Strongly deviating values are not shown in the diagram.








Strongly deviating values are not correctly shown in the diagram.




PT 7/23		dichloroacetic acid - 3			
assigned value [ $\mu\text{g/l}$ ]*		45,83 $\pm$ 6,54			
upper tolerance limit [ $\mu\text{g/l}$ ]		72,4			
lower tolerance limit [ $\mu\text{g/l}$ ]		25,12			
lab code	result [ $\mu\text{g/l}$ ]	$\pm$	z-score	Z <sub>U</sub> -score	assessm.**
1	51,8	10	1,0	0,4	s
4	52			0,5	s
5	62,1	24	1,3	1,2	s
8	58,1	13	1,7	0,9	s
9	32,6	6,5	-2,9	-1,3	s
12	30,9			-1,4	s
13	3,85			-4,1	u
14	24,8			-2,0	s
15	14,5			-3,0	u
16	40,3			-0,5	s
17	55,4	4,91	2,3	0,7	s
18	33,5	3,35	-3,4	-1,2	s
19	38,308			-0,7	s
20	51,5			0,4	s
21	33,4	16,7	-1,4	-1,2	s
22	17,6	0,93	-8,5	-2,7	q
23	45,1	6,77	-0,2	-0,1	s
26	54,3			0,6	s
27	43,1	4,96	-0,7	-0,3	s
28	53,6	7,4	1,6	0,6	s
29	41,2			-0,4	s
30	55,5	13,9	1,3	0,7	s
31	50,9	7,14	1,0	0,4	s
35	53,4			0,6	s
37	47,16			0,1	s
39	50,63			0,4	s
40	68,9	26,2	1,7	1,7	s
41	55,8			0,8	s
43	42,9	8,6	-0,5	-0,3	s
46	63,11			1,3	s
47	44,6	8,92	-0,2	-0,1	s



\* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor  $k=2$  corresponding to a confidence level of about 95%

\*\* s = satisfactory, q = questionable, u = unsatisfactory



Strongly deviating values are not shown in the diagram.







PT 7/23		trichloroacetic acid - 1			
assigned value [ $\mu\text{g/l}$ ]*		4,756 $\pm$ 0,599			
upper tolerance limit [ $\mu\text{g/l}$ ]		7,514			
lower tolerance limit [ $\mu\text{g/l}$ ]		2,607			
lab code	result [ $\mu\text{g/l}$ ]	$\pm$	z-score	Z <sub>U</sub> -score	assessm.**
1	4,96	1	0,3	0,1	s
4	5,35			0,4	s
5	5,47	1,96	0,7	0,5	s
8	5,2	1,2	0,7	0,3	s
9	3,65	0,73	-2,3	-1,0	s
12	3,78			-0,9	s
13	0,343			-4,1	u
14	2,5			-2,1	q
16	4,86			0,1	s
17	5,44	1,11	1,1	0,5	s
18	2,73	0,409	-5,6	-1,9	s
19	14,007			6,7	u
20	4,75			0,0	s
21	5,53	2,77	0,5	0,6	s
23	7,3	1,46	3,2	1,8	s
26	4,61			-0,1	s
27	3,13	0,332	-4,7	-1,5	s
28	5,02	1,36	0,4	0,2	s
29	4			-0,7	s
30	5,6	1,4	1,1	0,6	s
31	5,58	0,24	2,6	0,6	s
35	7,79			2,2	q
37	4,32			-0,4	s
40	10,1	4,55	2,3	3,9	u
41	5,58			0,6	s
43	4,73	0,95	0,0	0,0	s
46	10,66			4,3	u
47	5,24	1,05	0,8	0,4	s


\* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor  $k=2$  corresponding to a confidence level of about 95%

\*\* s = satisfactory, q = questionable, u = unsatisfactory




Strongly deviating values are not shown in the diagram.





Strongly deviating values are not correctly shown in the diagram.




PT 7/23		trichloroacetic acid - 2			
assigned value [ $\mu\text{g/l}$ ]*		19,89 $\pm$ 2,36			
upper tolerance limit [ $\mu\text{g/l}$ ]		31,42			
lower tolerance limit [ $\mu\text{g/l}$ ]		10,9			
lab code	result [ $\mu\text{g/l}$ ]	$\pm$	z-score	Z <sub>U</sub> -score	assessm.**
1	20,4	4,1	0,2	0,1	s
4	21			0,2	s
5	23,2	9,8	0,7	0,6	s
8	22,1	5	0,8	0,4	s
9	14,6	2,9	-2,8	-1,2	s
12	13,5			-1,4	s
13	1,5			-4,1	u
14	9,9			-2,2	q
15	5,26			-3,3	u
16	18,1			-0,4	s
17	22,4	3,44	1,2	0,4	s
18	15,5	2,33	-2,6	-1,0	s
19	54,845			6,1	u
20	19,8			0,0	s
21	17,1	8,55	-0,6	-0,6	s
23	26,1	5,2	2,2	1,1	s
26	20,3			0,1	s
27	16,2	1,72	-2,5	-0,8	s
28	18,7	4,74	-0,4	-0,3	s
29	20,6			0,1	s
30	23,3	5,83	1,1	0,6	s
31	22,4	0,95	2,0	0,4	s
35	23,8			0,7	s
37	17,62			-0,5	s
39	23,74			0,7	s
40	39,9	18	2,2	3,5	u
41	23,3			0,6	s
43	23,2	4,6	1,3	0,6	s
46	25,08			0,9	s
47	21,6	4,33	0,7	0,3	s


\* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor  $k=2$  corresponding to a confidence level of about 95%


\*\* s = satisfactory, q = questionable, u = unsatisfactory






Strongly deviating values are not shown in the diagram.






Strongly deviating values are not correctly shown in the diagram.




PT 7/23		trichloroacetic acid - 3			
assigned value [ $\mu\text{g/l}$ ]*		49,2 $\pm$ 4,88			
upper tolerance limit [ $\mu\text{g/l}$ ]		72,5			
lower tolerance limit [ $\mu\text{g/l}$ ]		30,29			
lab code	result [ $\mu\text{g/l}$ ]	$\pm$	z-score	Z <sub>U</sub> -score	assessm.**
1	49	9,8	0,0	0,0	s
4	53,3			0,4	s
5	52,5	18,5	0,3	0,3	s
8	51,9	11,8	0,4	0,2	s
9	30,5	6,1	-4,8	-2,0	s
12	29,3			-2,1	q
13	2,35			-5,0	u
14	26,3			-2,4	q
15	11,9			-3,9	u
16	44,4			-0,5	s
17	53,2	4,74	1,2	0,3	s
18	61,6	9,24	2,4	1,1	s
19	65,734			1,4	s
20	43,1			-0,6	s
21	43,3	21,7	-0,5	-0,6	s
23	43,8	8,76	-1,1	-0,6	s
26	50,9			0,1	s
27	41,4	4,39	-2,4	-0,8	s
28	48,5	9,05	-0,1	-0,1	s
29	41			-0,9	s
30	57,3	14,3	1,1	0,7	s
31	55,1	2,33	2,2	0,5	s
35	53,3			0,4	s
37	60,24			0,9	s
39	57,37			0,7	s
40	95,4	42,9	2,1	4,0	u
41	53,6			0,4	s
43	50,3	10	0,2	0,1	s
46	57,22			0,7	s
47	53,2	10,6	0,7	0,3	s


\* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor  $k=2$  corresponding to a confidence level of about 95%

\*\* s = satisfactory, q = questionable, u = unsatisfactory




Strongly deviating values are not shown in the diagram.






Strongly deviating values are not correctly shown in the diagram.




PT 7/23		dibromoacetic acid - 1			
assigned value [ $\mu\text{g/l}$ ]*		4,986 $\pm$ 0,699			
upper tolerance limit [ $\mu\text{g/l}$ ]		7,876			
lower tolerance limit [ $\mu\text{g/l}$ ]		2,733			
lab code	result [ $\mu\text{g/l}$ ]	$\pm$	$\zeta$ -score	$z_U$ -score	assessm.**
1	5,61	1,1	1,0	0,4	s
4	4,48			-0,4	s
5	5,48	1,68	0,5	0,3	s
8	5,4	1,1	0,6	0,3	s
9	5,52	1,1	0,8	0,4	s
12	5,32			0,2	s
13	0,849			-3,7	u
14	2,6			-2,1	q
15	1,33			-3,2	u
17	5,61	0,434	1,5	0,4	s
18	2,2	0,991	-4,6	-2,5	q
19	4,927			-0,1	s
20	5,03			0,0	s
21	4,79	2,4	-0,2	-0,2	s
22	2,57	0,12	-6,8	-2,1	q
23	6,81	1,022	2,9	1,3	s
26	6,52			1,1	s
27	1,74			-2,9	q
28	5,33	0,6	0,7	0,2	s
29	2,68			-2,0	s
30	5,5	1,38	0,7	0,4	s
31	6,15	1,05	1,8	0,8	s
35	7,03			1,4	s
37	5,74			0,5	s
39	5,96			0,7	s
40	7,76	2,25	2,4	1,9	s
41	6,09			0,8	s
43	3,5	0,7	-3,0	-1,3	s
46	7,91			2,0	s
47	7,24	1,45	2,8	1,6	s



\* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor  $k=2$  corresponding to a confidence level of about 95%

\*\* s = satisfactory, q = questionable, u = unsatisfactory



Strongly deviating values are not shown in the diagram.









PT 7/23		dibromoacetic acid - 3			
assigned value [ $\mu\text{g/l}$ ]*		46,43 $\pm$ 6,97			
upper tolerance limit [ $\mu\text{g/l}$ ]		73,36			
lower tolerance limit [ $\mu\text{g/l}$ ]		25,45			
lab code	result [ $\mu\text{g/l}$ ]	$\pm$	$\zeta$ -score	$z_U$ -score	assessm.**
1	51,9	10	0,9	0,4	s
4	49,3			0,2	s
5	57,6	15,2	1,3	0,8	s
8	47,2	10	0,1	0,1	s
9	42,2	8,4	-0,8	-0,4	s
12	48,4			0,1	s
13	8,99			-3,6	u
14	26,5			-1,9	s
15	15,3			-3,0	u
16	41,3			-0,5	s
17	52,9	8,13	1,2	0,5	s
18	48,1	21,6	0,1	0,1	s
19	28,447			-1,7	s
20	49,5			0,2	s
21	45,3	22,7	-0,1	-0,1	s
22	26,1	0,79	-5,8	-1,9	s
23	43,1	6,465	-0,7	-0,3	s
26	64,5			1,3	s
27	29,2	3,65	-4,4	-1,6	s
28	55,1	11,6	1,3	0,6	s
29	24			-2,1	q
30	55,1	13,8	1,1	0,6	s
31	62,1	10,6	2,5	1,2	s
35	62,3			1,2	s
37	48,18			0,1	s
39	54,5			0,6	s
40	75,6	21,92	2,5	2,2	q
41	57,1			0,8	s
43	37	7,4	-1,9	-0,9	s
46	66,2			1,5	s
47	65,8	13,2	2,6	1,4	s



\* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor  $k=2$  corresponding to a confidence level of about 95%

\*\* s = satisfactory, q = questionable, u = unsatisfactory



Strongly deviating values are not shown in the diagram.







PT 7/23		dibromoacetic acid - 2			
assigned value [ $\mu\text{g/l}$ ]*		19 $\pm$ 2,59			
upper tolerance limit [ $\mu\text{g/l}$ ]		30,01			
lower tolerance limit [ $\mu\text{g/l}$ ]		10,41			
lab code	result [ $\mu\text{g/l}$ ]	$\pm$	$\zeta$ -score	$z_U$ -score	assessm.**
1	22,4	4,5	1,3	0,6	s
4	18,8			0,0	s
5	23	6,76	1,1	0,7	s
8	23,2	4,9	1,5	0,8	s
9	17,8	3,6	-0,5	-0,3	s
12	20,6			0,3	s
13	3,87			-3,5	u
14	10,5			-2,0	s
15	6,17			-3,0	u
16	19,5			0,1	s
17	21,3	2,17	1,4	0,4	s
18	9,95	4,48	-3,5	-2,1	q
19	11,456			-1,8	s
20	20			0,2	s
21	18,5	9,25	-0,1	-0,1	s
22	9,92	0,3	-7,0	-2,1	q
23	24,9	3,735	2,6	1,1	s
26	25,2			1,1	s
27	6,32	0,79	-9,4	-3,0	u
28	21,1	0,29	1,6	0,4	s
29	10,3			-2,0	s
30	22,8	5,7	1,2	0,7	s
31	23	3,92	1,7	0,7	s
35	23,1			0,7	s
37	20,95			0,4	s
39	20,57			0,3	s
40	30,9	8,96	2,6	2,2	q
41	23,3			0,8	s
43	15,7	3,1	-1,6	-0,8	s
46	27,43			1,5	s
47	31,1	6,19	3,6	2,2	q



\* The stated uncertainty of the assigned value is the expanded uncertainty with a coverage factor  $k=2$  corresponding to a confidence level of about 95%

\*\* s = satisfactory, q = questionable, u = unsatisfactory



Strongly deviating values are not shown in the diagram.





Strongly deviating values are not correctly shown in the diagram.