



University of Stuttgart
Germany

University of Stuttgart • ISWA • Bandtäle 2 • 70569 Stuttgart
AQS Baden-Württemberg

To the participants of AQS Baden-Württemberg

**Institute for Sanitary Engineering,
Water Quality and Solid Waste
Management**

AQS Baden-Württemberg

Contact person

Dr. Frank Baumeister, Dr. Michael Koch,
Dipl.-Biol. Biljana Marić, Mirela Kordić

Contact details

Bandtäle 2
70569 Stuttgart
GERMANY
T +49 711 685-65446
F +49 711 685-53769
info@aqsbw.de
www.aqsbw.de

**Proficiency test 6/25
TW O6 – PFAS in drinking water**

2025-03-10

Dear Madam or Sir,

in May 2025 the execution of the above mentioned proficiency test (PT)
round "PFAS in drinking water" is planned.

Details about the PT round are enclosed. Please read them with care.

The PT will be executed according to the recommendations of the German
Federal Environment Agency from December 2003. These
recommendations "for the execution of PTs for the measurement of
chemical parameter and indicator parameter for the external quality
control of drinking water laboratories" (Bundesgesundheitsblatt 46 (12),
1094-1095) require, that drinking water laboratories must demonstrate
their competence for all parameters they are accredited for or they want to
be accredited for by a successful participation in a PT round within a cycle
of 2-3 years.

If you are interested in participation, please register online via our PT
portal. You will reach this portal via our website <http://www.aqsbw.de/en>.
For the first usage of the web portal, you must create a new user. If the e-
mail address is not known to the system, this user must be verified. Please
plan time for this registration process.

Application deadline: 28 March 2025

Bank

Baden-Württembergische
Bank Stuttgart – BW-Bank

IBAN

DE51 6005 0101 7871 5216 87

SWIFT/BIC

SOLADEST600

VAT-No.

DE147794196



If you have any questions, please do not hesitate to contact us:

AQS Baden-Württemberg, Bandtäle 2, 70569 Stuttgart, Germany


Phone: +49 711 685 65446

Telefax: +49 711 685 53769

E-Mail: info@aqsbw.de

Contact: Mirela Kordić, Biljana Marić, Dr. Frank Baumeister, Dr. Michael Koch

Best regards



Dr.-Ing. Michael Koch
Scientific director AQS

Dr.-Ing. Frank Baumeister
PT coordinator

Dipl.-Biol. Biljana Marić
assist. PT coordinator

Annex:

Details of the proficiency test exercise



**Details of the proficiency test round 6/25
TW O6 – PFAS in drinking water – (May 2025)**

Parameters

parameter	CAS-number of the parameter
Perfluorobutanoic acid (PFBA)	375-22-4
Perfluoropentanoic acid (PFPeA)	2706-90-3
Perfluorohexanoic acid (PFHxA)	307-24-4
Perfluoroheptanoic acid (PFHpA)	375-85-9
Perfluorooctanoic acid (PFOA)	335-67-1
Perfluorononanoic acid (PFNA)	375-95-1
Perfluorodecanoic acid (PFDA)	335-76-2
Perfluoroundecanoic acid (PFUnDA)	2058-94-8
Perfluorododecanoic acid (PFDoDA)	307-55-1
Perfluorotridecanoic acid (PFTrDA)	72629-94-8
Perfluorobutane sulfonic acid (PFBS)	375-73-5
Perfluoropentane sulfonic acid (PFPeS)	630402-22-1
Perfluorohexane sulfonic acid (PFHxS)	355-46-4
Perfluoroheptane sulfonic acid (PFHpS)	375-92-8
Perfluorooctane sulfonic acid (PFOS)	1763-23-1
Perfluorononane sulfonic acid (PFNS)	98789-57-2
Perfluorodecane sulfonic acid (PFDS)	335-77-3
Perfluoroundecane sulfonic acid (PFUnDS)	749786-16-1
Perfluorododecane sulfonic acid (PFDoDS)	79780-39-5
Perfluorotridecane sulfonic acid (PFTrDS)	791563-89-8
Sum of PFAS-4	
Sum of PFAS-20	

Matrix

Drinking water

Dates and deadlines

Registration deadline: 28 March 2025

Please register for this PT only via our PT portal (accessible via our website).

You will receive a confirmation of receipt by e-mail and your registration is documented in our PT portal.

Dispatch of the samples: 26 May 2025

**Deadline for submission of results 16 June 2025; 24:00h online via internet.
Results submitted after the deadline will not be accepted.**

Sample dispatch

Samples will be sent by courier service.

Sample details

- 3 Samples for the determination of PFAS in 2 x 250 ml plastic bottles each.

Permitted analytical methods



Limit of quantification

The laboratories have to ensure that they use methods with a limit of quantification of 1,5 ng/l for all PFAS.

Execution of the analysis

The samples must be analysed in the own laboratory with own personnel and own equipment. Subcontracting of the analysis is not allowed.

Evaluation and assessment of the single values

The statistical evaluation will be executed according to DIN 38402 – A45 or ISO/TS 20612 “Interlaboratory comparison for proficiency testing of analytical chemistry laboratories” with the combined estimator Hampel/Q-method, a method of robust statistics. The assigned value x_{pt} derived from the weighings of the spiked samples and the matrix content^{1,2} will be preferably used for the assessment of the single values. Only if this is not possible, the Hampel estimator as robust mean value of the participants’ data will be used.

If possible, the standard deviation for proficiency assessment σ_{pt} will be taken from the variance function for the calculation of the z_U -scores according to DIN 38402 - A45 (chapter 10.4) or ISO/TS 20612 respectively. σ_{pt} will be limited for all PFAS as follows:

- lower limit: 5 %
- upper limit: 25 %

A z-score for a result x is calculated for each measurement result derived from the assigned value x_{pt} and the standard deviation for proficiency assessment σ_{pt} :

$$z = \frac{x - x_{pt}}{\sigma_{pt}}$$

The z-score will be modified to a z_U -score with a correction factor for proficiency assessment (as described in the above mentioned standards).

The tolerance limits are defined as $|z_U|=2$.

The single results will be assessed as follows:

$ z_U \leq 2.0$	satisfactory
$2.0 < z_U < 3.0$	questionable
$ z_U \geq 3.0$	unsatisfactory

¹ Rienitz, O., Schiel, D., Güttler, B., Koch, M., Borchers, U.: A convenient and economic approach to achieve SI-traceable reference values to be used in drinking-water interlaboratory comparisons. *Accred Qual Assur* (2007) 12: 615-622.

² Koch, M., Baumeister, F.: Traceable reference values for routine drinking water proficiency testing: first experiences. *Accred Qual Assur* (2008) 13: 77-82.

Overall assessment

There is no overall assessment of the proficiency test round, but the single parameters are assessed.

A parameter is assessed as successful, if more than half of the values are assessed as "satisfactory".

In addition those values are assessed as "unsatisfactory":

- 1) that are not within the tolerance limit,
- 2) that are not determined (if the other samples of this parameters are analysed),
- 3) that are indicated with "lower than limit of quantification",
- 4) that have been subcontracted.

Participation fee

The participation fee will be 475 € plus transport costs.